

## Indira Gandhi Centre for Human Ecology, Environmental & Population Studies University of Rajasthan, Jaipur-302004

## इन्दिरा गांधी मानव पारिस्थितिकी, पर्यावरण एवं जनसंख्या अध्ययन केन्द्र राजस्थान विश्वविद्यालय, जयपुर -302004

DIRECTOR निदेशक

No./IGC/2023/24 7 7 दिनांक : 16/03/2023

## **Tender Notice Inviting Quotations**

Sealed quotation with terms and conditions are invited for the purchase of item (from reputed companies/ MNC) mentioned below in Indira Gandhi Centre for HEEPS, University of Rajasthan, Jaipur

| <ol> <li>High Precision Water Bath</li> <li>Specifications:         <ul> <li>Hi precision water bathe should have made of double walled construction.</li> <li>Inner chamber and Outer body made of thick stainless steel grade 304, painted with powder coating with concentric rings.</li> <li>Should have Microprocessor based digital PID Temperature Controller cum indicator.</li> <li>Temperature Range: Ambient +5Deg.C to 95 Deg.C with accuracy of ± 0.1 Deg.C.</li> <li>Fitted with stirrer to make even temperature &amp; high-quality element.</li> </ul> </li> </ol> | Description :  | Quantity | Approximately Amount (Including all taxes |
|--|--|----------|---|
| <ul> <li>Hi precision water bathe should have made of double walled construction.</li> <li>Inner chamber and Outer body made of thick stainless steel grade 304, painted with powder coating with concentric rings.</li> <li>Should have Microprocessor based digital PID Temperature Controller cum indicator.</li> <li>Temperature Range: Ambient +5Deg.C to 95 Deg.C with accuracy of ± 0.1 Deg.C.</li> </ul>   | 1. High Precision Water Bath   | 01       | 50000.00                                  |
| <ul> <li>Inner chamber and Outer body made of thick stainless steel grade 304, painted with powder coating with concentric rings.</li> <li>Should have Microprocessor based digital PID Temperature Controller cum indicator.</li> <li>Temperature Range: Ambient +5Deg.C to 95 Deg.C with accuracy of ± 0.1 Deg.C.</li> </ul>   | Specifications :   |          |   |
| <ul> <li>painted with powder coating with concentric rings.</li> <li>Should have Microprocessor based digital PID Temperature Controller cum indicator.</li> <li>Temperature Range: Ambient +5Deg.C to 95 Deg.C with accuracy of ± 0.1 Deg.C.</li> </ul>   | Hi precision water bathe should have made of double walled construction.                     |          |   |
| <ul> <li>Should have Microprocessor based digital PID Temperature Controller cum indicator.</li> <li>Temperature Range: Ambient +5Deg.C to 95 Deg.C with accuracy of ± 0.1 Deg.C.</li> </ul>   | • Inner chamber and Outer body made of thick stainless steel grade 304,                      |          |   |
| <ul> <li>cum indicator.</li> <li>Temperature Range : Ambient +5Deg.C to 95 Deg.C with accuracy of ± 0.1</li> <li>Deg.C.</li> </ul>   | painted with powder coating with concentric rings.   |          |   |
| Temperature Range : Ambient +5Deg.C to 95 Deg.C with accuracy of ± 0.1  Deg.C.   | Should have Microprocessor based digital PID Temperature Controller                          | 1        |   |
| Deg.C.   | cum indicator.   |          |   |
|  | • Temperature Range : Ambient +5Deg.C to 95 Deg.C with accuracy of $\pm$ 0.1                 |          |   |
| Fitted with stirrer to make even temperature & high-quality element.   | Deg.C.   |          |   |
|  | <ul> <li>Fitted with stirrer to make even temperature &amp; high-quality element.</li> </ul> |          |   |
| • Size: 430 x 230 x 265 mm.  | • Size: 430 x 230 x 265 mm.  |          |   |
|  |  |          |   |

Firms are requested to give details of the specification, delivery, and availability. Quotations should reach the office of the Director on or before March 20, 2023 (4:30 p.m.). Quotations will be opened on March 21, 2023 at 01.30 PM.

## **Terms and Conditions:**

- 1. Quotations received after due date will not be considered in any case.
- 2. Please mention the rate for C.IF/F.O.R. Jaipur inclusive of all charges and taxes.

**DIRECTOR**