



**Centre for Non Conventional Energy Resources  
University of Rajasthan  
Jaipur-302004**

Ref. No. CNCER/2021/63

Dated: 23-02-2021

**Corrigendum to the Tender**

The dimensions of the quoted sample holder with heater for homemade siever type hydrogen storage setup are 95 mm (H) x 35 mm (OD) x 28 mm (ID). It should be fabricated by very high quality Stainless Steel (SS). The temperature range for heater should be from Room Temperature to 500<sup>0</sup> C.

It should be read with the Tender No. CNCER/2021/62 dated 22 February 2021.

Director

C.N.C.E.R.

**Director  
CNCER**

University of Rajasthan, JAIPUR